



## Conductivity Tester Instructions:

### Using the Hanna Instruments TDS 1 Meter

#### Step 1

Collect a water sample from your drinking fountain, faucet, or hose bib. Fill the container **¾ full**.



#### Step 2

Pull the black cap off.



#### Step 3

Turn the conductivity tester on by sliding the black switch to the left. The window will read **000**.



#### Step 4

Lower the conductivity tester in the water sample container up to the MAX LEVEL LINE.

Make sure the tester does not touch the bottom of the container. The reading will appear in the window.



#### Step 5

Call your water retailer if the number is greater than 500. Water retailer contact information:

San Jose Municipal Water	(408) 277-4036
City of Milpitas	(408) 586-3100
San Jose Water Company	(408) 279-7900
City of Santa Clara	(408) 615-2000

This publication can be made available upon request in alternative formats such as Braille, large print, audio tape or computer disc. Requests can be made by calling (408) 535-8550 (V), or (800) 735-2929 (CRS). For Spanish version of these instructions or questions and concerns about using the conductivity tester, call Tina Simmons, South Bay Water Recycling Customer Services Specialist, at (408) 363-4717.